



Ref. Certif. No.

US-17017-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Switch Mode Power Supply

Name and address of the applicant
Nom et adresse du demandeur

RAM TECHNOLOGIES L L C
UNIT 12
29 SOUNDVIEW RD
GUILFORD CT 06437, USA
RAM TECHNOLOGIES L L C
UNIT 12
29 SOUNDVIEW RD
GUILFORD CT 06437, USA
See Page 2

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{eme} page
Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

See Page 2

Trademark (if any)
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur
Model / Type Ref.
Ref. De type

PFC600PCX

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{eme} page

This report comprises 2 enclosures.

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

10ME09619

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pflingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Signature:

Date: 2011-05-05

Jolanta M. Wroblewska



Ref. Certif. No.

US-17017-UL

Factories:

RAM TECHNOLOGIES L L C
UNIT 12
29 SOUNDVIEW RD
GUILFORD CT 06437, USA

Ratings:

Input Ratings:

100-240 Vac, 50-60 Hz, 7.5 A max

Output Ratings:

550W Continuous / 600W Peak

Max output current: +3.3 Vdc, 24 A; +5 Vdc, 24 A; +12 Vdc, 45.8

A/50A Peak; -12 Vdc, 0.5 A; 5 Vdc(5VSB), 3A/3.5AP;

Total output of +12 V, +5 V, +3.3 V, -12 V & 5 VSB, shall not exceed

550 Watts Continuous / 600W peak

Max Output from 3.3V and 5V Output shall not exceed 140 Watts.

Additional information (if necessary)

Information complémentaire (si nécessaire)



Date: 2011-05-05

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Signature:

Jolanta M. Wroblewska